

1565-B Series

1565-B汎用サウンドレベルメータはOSHAで要求される測定及び一般的なノイズの調査に理想的な計器です。この人気のあるメータの重さは1ポンド未満で、単純なプッシュボタン操作、バッテリー駆動、セラミックマイクを備えています。1565-Bは単体でも購入できますが、音響キャリブレータ、ウインドスクリーン、キャリング/保管用ケースも備えております。

- 40 to 140 dB
- ANSI Type 2
- A, B, C weighting
- Fast/slow meter response
- MESA approved



Model 1565-B Sound-Level Meter

仕様

サウンドレベル: 40 1 ~ 140 dB re 20 µN/m²

重み付け: A, B, and C. Conforms to ANSI S1.4-1971 Type 2 and IEC651.

メータ: rms response with fast and slow speeds.

入力: マイクロフォン: Lead-zirconate-titanate ceramic.

出力: > 1.2 Vrms behind 620 Ω with meter at full scale; will drive oscilloscope or low-impedance headphones. HARMONIC DISTORTION: < 0.5% (0.1% typical) from 32 Hz to 8 kHz, C-weighted with meter at full scale. Any load impedance may be connected.

校正: Can be pressure calibrated at 125, 250, 500, 1000, and 2000 Hz with GR 1562 Calibrator, and at 1000 Hz with GR 1987*.

環境: TEMPERATURE: -10 to 50°C operating; -40 to +60°C storage, with batteries removed; coefficient of sensitivity ~ +0.02 dB/°C at 6 dB below full-scale meter reading. HUMIDITY: 0-90% RH within 0.5 dB. MAGNETIC FIELD: 1 oersted (80 A/m) 50- or 60- Hz field causes about 45 dB C-weighted (negligible A-weighted) indication when meter is oriented for maximum sensitivity to field. A

1-oersted, 400-Hz field causes about 58 db A-weighted indication when meter is oriented for maximum sensitivity. VIBRATION: With the microphone attached, the highest meter indications (without interference from background noise) when the instrument is vibrated at an acceleration of 0.1 g, are 67 dBA at 63 Hz, 68 dBA at 250 Hz, and 70 dBA at 1000 Hz. When the microphone is replaced by an equivalent impedance, the same test gives no meter indication.

付属品: Microphone safety cap, instruction manual, carrying case.

電源: Two 9V alkaline batteries (NEDA 1604AC) supplied, provide 50 hours of operation.

形状: シールド付プラスティックケース

大きさ: : 16.5 cm H x 0.92 cm W x 0.53 cm D (6.5" x 3.63" x 2.09").

重さ: 0.48 kg (13 oz) net, 1.4 kg (3 lb) shipping.

* At 125, 250, 500, 1000, 2000, and 4000 Hz with GR 1986 Calibrator.

オーダー情報

1565-9702 1565-B Sound-Level Meter

1562-9600 Carrying Case



IET LABS, INC. in the GenRad Tradition
534 Main Street, Westbury, NY 11590

Electronic cat/1482/09-15-03

www.ietlabs.com

TEL: (516) 334-5959 • (800) 899-8438 • FAX: (516) 334-5988